


Supplement

| | | | |
|----------------|------------|-------------------|------|
| Title: | 9040 Users | Supplement Issue: | 1 |
| Part Number: | 2438546 | Issue Date: | 6/13 |
| Print Date: | April 2005 | Page Count: | 2 |
| Revision/Date: | 2, 5/11 | | |

This supplement contains information necessary to ensure the accuracy of the above Instruction Sheet.

Change #1, 65800

On page 5, add the following to the **Symbols** table:

| | |
|---|--|
|  | Conforms to relevant South Korean EMC Standards. |
|---|--|

On page 10, remove **Pollution Degree 2**, **Type of Protection**, **Safety Specifications**, **Electrical Safety**, **Maximum Operating Voltage**, **Protection Level** and **Nominal Voltage** sections and add:

| | |
|----------------|--|
| Altitude..... | 2000 m |
| Humidity..... | 15 % to 80 % |
| Safety..... | IEC 61010-1: 600 V CAT III / 300 V CAT IV, Pollution Degree 2 |
| IP Rating..... | IEC 60529: IP40 |

Maximum Operating Voltage (U_{me})

600 V

Nominal Voltage

40 to 600 V ac

Electromagnetic Environment IEC 61326-1: Industrial

Electromagnetic Compatibility Applies to use in Korea only. Class A
Equipment (Industrial Broadcasting &
Communication Equipment)^[1]

- [1] This product meets requirements for industrial (Class A) electromagnetic wave equipment and the seller or user should take notice of it. This equipment is intended for use in business environments and is not to be used in homes.